Displacement and Proximity

09-1

Types of Displacement Transducers

09-2

 $Displacement\ {\it transducers}\ {\it measure}\ {\it the}\ {\it location}\ {\it of}\ {\it an}\ {\it object}.$

Proximity transducers determine when an object is near.

Criteria Used in Selection of Transducer

 \bullet How much force can transducer exert on object?

Requirement can be as low as zero when the object is small or cannot be touched.

- What is the range of positions to be measured?
- How fast will the object be moving?
- What accuracy is required?

Potentiometer.

09-2

Object moves tap on a variable resistor.

Linear Variable Differential Transformer (LVDT).

Object moves core of three-winding inductor. $\,$

Capacitive Transducers.

Object is either one plate of a capacitor, or moves the capacitor's dielectric.

Coded

Object moves surface covered with code marks. A transducer reads the code marks.

There are many other ways to measure position \dots

- ... these will not be covered individually ...
- ... but may be explained as part of problems.

09-1

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09.

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from lsli09

09-2

09-4

09-3

Potentiometers

09-3

09-1



Construction and Operation:

Resistive material formed into either a strip or coil.

An external object, through a mechanical linkage, moves the tap.

Tap slides along resistive material.

Tap and both ends of resistive material connected to leads.

Miscellaneous Facts:

These are built for instrumentation purposes. (Unlike the volume control on a radio.)

Multiple-turn potentiometers can be made precise. (Percent calibration error less than 1%.)

09-4

09-2

Desirable Characteristics

- \bullet Easy to read output.
- Linear response. Other responses also possible.
- \bullet In expensive.

Undesirable Characteristics

- Adds friction to motion of object.
- Can wear out.
- Limited precision.
- Can only measure relatively large displacements.

Capacitive Displacement Transducers

Symbol: no special symbol used.

Two types discussed: $moving\ plate$ and $moving\ dielectric$.

Construction and Principle of Moving Plate Type:

One plate of capacitor mounted to a fixed surface.

The other plate mounted to the object.

Capacitance changes with position of object.

Possible ways of mounting and moving plates:

- Plates move normal to plane of plates.
- Plates move parallel to plane of plates.
- Plates, shaped like pie slices, rotate.

Construction and Principle of Moving Dielectric Type:

Both plates of capacitor mounted to a fixed surface.

The dielectric mounted to an object.

Capacitance changes with position of object.

Possible ways of moving dielectric:

- Dielectric slides.
- \bullet Dielectric is a fluid that flows in and out of the capacitor.

09-5

09-6

This varies for each type.

Moving-Plate Type

Model Function

If the area of the plates is large compared to the distance between them

and one plate moves normal to its plane:

$$H_{\rm t1}(x) = C\frac{\rm m}{x}, \quad x>0\,{\rm m}. \label{eq:Ht1}$$

Moving-Dielectric Type

$$H_{t1}(x) = C_0(1 + Kx),$$

where K is a constant determined by the transducer construction.

There are many variations on capacitive sensors, so there are many model functions.

Derivation and use of model functions for capacitive sensors is beyond the scope of the course. $\,$

09-5

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from 18li09.

09-5

09-7

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from lsli09

09-6

09-8

09-6

09-7

Typical Conditioning Circuits

Oscillator

Capacitive transducer may be part of an LC pair in an oscillator.

Capacitance measured by counting frequency.

Bridge

Capacitive transducers can be placed in a bridge. (In the same way as resistive transducers.)

AC voltage placed across bridge \dots

... instead of DC, as with resistive transducers.

Bridge voltage is amplified, rectified, then measured.

Analysis of these circuits is beyond the scope of this course.

09-8

09-6

Desirable Characteristics

- No physical contact with object. (Moving dielectric type.)
- No wear with motion.
- Can measure small displacements. (Including the position of a diaphragm in some microphones.)


Undesirable Characteristics

- Can only measure small displacements.
- Conditioning circuits might be more troublesome.

Linear Variable Differential Transformer (LVDT)



Construction:



A three-winding transformer with a movable core.

Primary winding runs length of transformer.

Each secondary covers one half of transformer.

Secondaries are connected to oppose each other.

Object is connected to core.

09-9

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from lsli09.


09-9

09-11

09-9

09-10 09-10

Principle of Operation



When core is centered . .

... voltage at two secondaries are equal.

Because they are connected in opposition . . .

- \dots voltages cancel \dots
- ... and so output is zero.

When core is off center \dots

- \dots voltage at one secondary \dots
- ... is higher than at other.

Output voltage is linearly related to core position.

09-10 EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09. 09-10

09-11

Model Function

 $H_{\mathrm{t1}}(x,t) = Kxv_p(t), \ldots$

... where K is a constant, $v_p(t)$ is the voltage at the primary winding, and t is time

Note, if x is negative then phase of output is reversed.

In Typical LVDTs:

The primary voltage can range from 1 to 10 V.

Frequency from $50\,\mathrm{Hz}$ to $25\,\mathrm{kHz}$ can be used for primary winding.

Desirable Characteristics

- Low wear.
- High speed, can measure vibrations.
- \bullet Can measure small displacements, $<0.1\,\mathrm{nm}.$ (Precision determined by conditioning circuit.)
- Low repeatability error.

Undesirable Characteristics

- \bullet Complex conditioning circuit needed.
- Expensive.

09-12

Coded Displacement Transducers (CDT)

09-12

These can be either:

- Relative.
 - Transducer indicates $\it change$ in position, not actual position.
- $\bullet \ Absolute.$


Transducer indicates actual position.

They can measure:

- Linear displacement. Position along a line.
- Angular displacement. Angular position.

A generic CDT will be described, then the various types.

Physical Construction:



A strip or disk of some material, such as plastic, is mounted on object or on something linked to the object.

Strip or disk has marks in one or more tracks or rings.

Transducers, which can read the marks, are mounted in a fixed position.

A variety of transducers can be used to measure the marks.

Output of transducers is converted into a digital form.

Symbols d_i will denote the digital form of the transducer output.


The strip or disk is usually in a sealed package, to keep dirt out.

CDTs differ in the number of tracks and how the marks are placed.

 ${\bf 09\text{-}14}$ Relative One-Way Angular Coded Displacement Transducers

Use: measurement of angle in systems limited to one-way rotation.

Marking



Two rings are used, position and index.

There are many marks on the position ring, regularly spaced.

Call signal from position ring $d_{\rm p}$.

There is a single mark on the index ring.


Call signal from the index ring d_i .

09-13 EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09.

09-14 EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09

09-14

09-15 09-15



A counter is used.

Output $d_{\rm p}$ increments the counter, and $d_{\rm i}$ resets the counter.

Further processing may be done to get the output in the desired form, for example radians.

Model Function

This includes both the CDT and the counter.

 $H_{\rm t1}(x)=rac{x}{2\pi\,{
m rad}}N,$ where N is the number of position-ring marks.

Note that the output is incorrect \dots

 \dots from power-on \dots

 \dots until the sensing of the index mark

09-16 Example of Relative One-Way Angular CDT 09-16

Design a system to convert process variable $x \in [0 \text{ rad}, 2\pi \text{ rad}]$, the rotation angle of wheel, to a floating point quantity, H(x) = x/ rad, to be written in variable theta. The wheel will rotate in only one direction. The precision must be at least 0.001 rad.

Solution

Use a relative one-way angular CDT.

Link the CDT disk to the wheel \dots

 \dots so that one turn of the wheel results in one turn of the CDT.

(It would also have been possible to put the marks directly on the wheel, but this might have been less convenient.)

Need at least 6284 marks.

Assume an N=10,000-mark CDT is available.

Then a $\lceil \log_2 N \rceil = \lceil \log_2 10000 \rceil = 14\text{-bit counter}$ is needed.

Using these, $H_{\rm t1}(x) = \frac{x}{2\pi \, {\rm rad}} 10000$.

Need $H_{\rm f}(H_{\rm t1}(x)) = H(x)$.

Let $r = H_{\rm t1}(x) = \frac{x}{2\pi \, {\rm rad}} N$; solving, $x = \frac{2\pi \, {\rm rad}}{N} r$.

Then $H_{\mathrm{f}}(r) = H\left(\frac{2\pi \operatorname{rad}}{N}r\right) = \frac{2\pi}{N}r.$


theta = r * 0.000628319;

Conditioning

09-17

Use: measurement of angle or linear displacement in systems in which displacement will frequently be equal to some value (e.g.,0).

Marking



Description is for linear CDT; angular CDT is similar.

Three tracks are used, position, direction, and index.

Call the signal from the position track $d_{\rm p},$ direction track $d_{\rm d},$ and index track $d_{\rm i}.$

There are many marks on the position and direction tracks, regularly spaced.

Position and direction tracks partially overlap.

When $d_{\rm p}$ makes a 0 to 1 transition, $d_{\rm d}$ gives the direction of travel.

There is a single mark on the index track.

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09.

Displacement

Position

Direction

Index

n-Bit Up/Down Counter

Reset

Conditioning Circuit

Up/Down

(0=Up, 1=Down)

An up/down counter is used.

Analog

Output $d_{\rm p}$ clocks the counter and $d_{\rm d}$ determines the direction of count.

Digital

Clock

As in the one-way CDT, d_i resets the counter.

Conditioning Circuit

Further processing may be done to get the output in the desired form, for example radians.


09-18 EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from 1sli09.

09-19 Absolute Binary-Coded Displacement Transducers

Use: measurement of linear or angular position.

 ${\bf Marking}$

09-17



For an n-bit precision device, n+1 tracks or rings are used.

One track is called the clock track, the others are called position tracks.

There are 2^n clock marks, spaced regularly.

Call signal from clock track d_ϕ and the signal from the i 'th position track $d_i.$

Let ϵ be the largest distance (in the direction of displacement) between transducers. (In an ideal system, $\epsilon=0$).

Marks are positioned so that at position $x \pm \epsilon$,

if there is a transition in d_{ϕ} then there is no transition in any d_{i} .

As one might expect, the d_i give the position in binary.


09-20

Conditioning

An n-bit register is used.

The d_i connect to the register's data inputs.


The register is clocked with d_{ϕ} .



09-19

09-20

Purpose of Clock Marks



Consider a system in which there were no clock marks.

Output would be conditioned transducer signals, d_i , rather than register output.

Only n tracks would be needed for 2^n distinct positions.

This won't work. Consider a change from position 7 to 8:

Before: $d = 0111_2$. After: $d = 1000_2$.

Either all mark-reading transducers change at the same instant,

or $d \in \{0000,0001,0010,0011,0100,0101,0110,1001,1010,1011,1100,1101,1110,1111\}$ in the moments between.

Why Binary?

09-22



Other (not 100% serious) possibilities:

- Binary-coded decimal.
- Seven-segment display. Output can be fed to seven-segment displays.
- \bullet ASCII numerals. (E.g., 123). Read the number into a computer without ever having to do a binary conversion.
- \bullet ASCII words. (E.g., One hundred twenty three.). It can be done, but who would buy it?

The cost of a CBT is proportional to the number of tracks.

Binary is good because the minimum number of tracks are used, <u>plus one.</u>

09-21

 $\mbox{EE 4770 Lecture Transparency.} \mbox{ Formatted} \mbox{ } 13{:}17, \mbox{ } 10 \mbox{ February 1999 from } 1{sli09}.$

09-21

09-23

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from lsli09

09-22

09-23 Absolute Gray-Coded Displacement Transducers

Use: measurement of linear or angular position.

Marking

For an n-bit precision device, n tracks or rings are used.

All tracks are called position tracks.

Call signal from the i'th position track d_i .

As one might expect, the d_i give the position in gray code.

Gray Code Properties

Gray-coded number: sequence of zeros and ones (like binary).

Let a be a nonnegative integer.

Gray-code for a differs from gray-code for a+1 in exactly one digit.

Therefore, with gray code, only 1 transducer's output will change at a time—no races.

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from lsli09.

09-24

09-22

09-24

Gray Code Definition

Gray code is defined for non-negative integers.

Let $B \in [0, 2^n)$. (That is, B is any number from 0 to $2^n - 1$.)

Let b_i be digit i in B's binary representation, for $i \in [0,n)$. (The least-significant digit is b_0 .)

Symbol g_i is digit i in B's gray-code representation if

$$g_{n-1}=b_{n-1} \quad \text{ and } \quad g_i=b_{i+1}\oplus b_i, \quad \text{ for } i\in [0,n-1).$$

Symbol b_i is digit i in B's binary representation if

$$b_{n-1} = g_{n-1}$$
 and $b_i = b_{i+1} \oplus g_i$, for $i \in [0, n-1)$.

09-25 09-25

Gray Code Examples

Four Bits

 $7~(\mathrm{decimal}) = 0111~(\mathrm{binary}) = 0100~(\mathrm{gray})$

8 (decimal) = 1000 (binary) = 1100 (gray)

Five Bits

 $11 \; ({\rm decimal}) = 01011 \; ({\rm binary}) = 01110 \; ({\rm gray})$

12 (decimal) = 01100 (binary) = 01010 (gray)

 $13~(\mathrm{decimal}) = 01101~(\mathrm{binary}) = 01011~(\mathrm{gray})$

14 (decimal) = 01110 (binary) = 01001 (gray)

Twenty Bits

123,456 (decimal)

=0001 1110 0010 0100 0000 (binary)


=0001 0001 0011 0110 0000 (gray).

09-26

Gray- and Binary-Coded Displacement Transducers

09-26

Four-Bit Binary and Gray:




Bit 2 Bit 1 Bit 0 Clock

Bit 3 (MSI Bit 2 Bit 1 Bit 0

Eight-Bit Binary:



Nine-Bit Gray:



09-25

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from lsli09.

09-25

09-27

09-26 EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09

09-26

09-27 Coded Displacement Transducers Characteristics

Desirable Characteristics

- Linear (or any other response needed).
- Accurate and precise.
 Strip or disk can have 65536 or more marks per track or ring.
- Low wear.
- \bullet Simple conditioning circuit.

Undesirable Characteristics

• Contact with object required, or object must have marks.

Form of Marks

Ink on a ceramic or glass substrate.

Holes in some material. (Read optically or mechanically.)

Current-conducting ink on an insulating substrate.

09-28

Proximity Transducers

09-28

Detect when an object is near.

Types of Transducers

- Reed switch.
 Switch closed by magnet on object.
- Hall effect.

Magnetic field from object induces a potential . . .

- \ldots on a block of semiconductor.
- \bullet Magnetic reluctance.

An object passes through a magnetic field...

... inducing a voltage in a coil.

Plus many more, not covered.

No conditioning circuits will be given for the proximity transducers.

 ${\bf Construction}$


and Operation

Magnetic Field

Magnetic

Object

Construction and Operation:



Small glass tube with two switch contacts.

Magnet pulls switch contacts closed.

As with all switches, it does not go from open to closed neatly. (When closing, it might go from open to closed many times before settling into a closed state.)

Conditioning Circuit

Output might have to be "debounced" (low-pass filtered).

Desirable Characteristic

• Cheap.

Undesirable Characteristics

- Low speed.
- Limited life.

09-29 EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09.

09-29

Consists of a block of semiconductor material. There is no PN junction.

The object detected must produce a magnetic field.

If it's not naturally magnetic and you can't glue a magnet to it, you can't use a hall-effect transducer.

Consider the three axes normal to the block's faces.

A constant current is passed along one axis.

The transducer is positioned so that the magnetic field from the object is parallel to a second axis.

And, voilà, a potential is induced on the faces normal to the third axis.

Contacts on these faces connect to leads.

The output of the transducer is the voltage.


09-30 EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from 1sli09. 09-30

09-31

Magnetic Reluctance Transducer

09-31

Construction and Operation



Consists of a coil of wire wrapped around a c-shaped magnetic core.

Magnetic field lines pass through gap in core.

Object passes through the gap.

This changes the strength of the magnetic field which. . .

induces a voltage in the coil ends.

Voltage is proportional to $\underline{\text{rate of change}}$ of field strength.

09-32

Other Proximity Transducers

09-32

09-30

Many ways of building these.

Type depends upon what is being detected and where it is being detected.

Broken light beam.

Radiated heat.

Sound.

:

09-31

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from Isli09.

09-32

09-31

EE 4770 Lecture Transparency. Formatted 13:17, 10 February 1999 from lsli09.